

EU Declaration of Conformity

We hereby declare under our sole responsibility that the EvalTech system meets the relevant provisions of the following European Union Directives:

- Council Directive 93/42/EEC of 14 June 1993 concerning medical devices as amended by Directive 2007/47/EC (MDD)
- **Directive 2006/42/EC** of the European Parliament and of the Council of 17 May 2006 on machinery as amended by Regulation (EU) 2019/1243
- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)
- **Directive 2014/53/EU** of the European Parliament and of the Council of 16 April 2014 on radio equipment (Radio Equipment Directive)

The EvalTech has undergone a conformity assessment procedure required by the MDD and is manufactured in harmony with the Technical Documentation compiled as defined in the relevant Directives and retained by BTE.

Product information in regard to the MDD and RoHS Directives:

Manufacturer	BTE Technologies 7455-L New Ridge Road Hanover, MD 21076, USA www.btetechnologies.com Telephone: 410.850.0333 Email: Service@btetechnologies.com.	
Product Identification	Device Trade Name: EvalTech Device Name: ER Model: EvalTech	
UDI-DI	10850390007328	
EMDN (CND) code	Z120616 - PHYSICAL THERAPY AND REHABILITATION SYSTEMS	
Intended Purpose	The EvalTech is intended to be used for musculoskeletal testing. Applications include occupational and physical therapy and industrial rehabilitation.	
Device Classification (MDD)	Class I	
Classification Rule (MDD)	Rule 12	



Route to Compliance (MDD)	Annex VII of the Medical Devices Directive	
Device Classification (MDR)	Class IIa	
Classification Rule (MDR)	Rule 11	
CE Marking Provision	Under Medical Device Regulation (EU) 2017/745 (MDR), the device will be up-classified to class IIa due to changed software classification rules. Based on the MDR Article 120 §3, the EvalTech can be placed on the EU market as a class I device until May 26, 2024 provided that the device	
	 will continue to comply with the MDD, there will be no significant changes in the design and intended purpose, and the device will comply with the MDR requirements for post market surveillance, vigilance, and registration of economic operators and of devices 	
Authorized Representative EC REP	Emergo Europe Prinsessegracht 20 2514 AP, The Hague The Netherlands Telephone: +31.70.345.8570 Emails: EmergoEurope@ul.com EmergoVigilance@ul.com	

The device is CE marked since 2008.

Signed for on behalf of BTE Technologies

Ewa Kaczanowska

PRRC/Regulatory Manager

Whenowsko

BTE Technologies

Hanover, MD

May 20, 2021



Addendum to the original Declaration of **Conformity**

Per 31 January 2023, the address of our EU Authorized Representative as listed on the original DoC has changed.

OLD ADDRESS AUTHORIZED REPRESENTATIVE				
Name of company	Address	Telephone/email		
Emergo Europe	Prinsessegracht 20	+31.70.345.8570 - phone		
	2514 AP The Hague	EmergoEurope@ul.com		
	The Netherlands			

NEW ADDRESS AUTHORIZED REPRESENTATIVE				
Name of company	Address	Telephone/email		
Emergo Europe	Westervoortsedijk 60	+31.70.345.8570 - phone		
,	6827 AT Arnhem	EmergoEurope@ul.com		
	The Netherlands			

COMPANY REPRESENTATIVE:

Mecronowska

Ewa Kaczanowska

Regulatory Manager/PRRC

June 2, 2023

PHONE 800.331.8845 | 410.850.0333

FAX 410.850.5244